



Particle Size Surface Area Pore Size Distribution Density

CE DECLARATION OF CONFORMITY

Micromeritics Instrument Corporation, Inc.
One Micromeritics Drive
Norcross, Georgia 30093-1877 U.S.A.
ISO-9001 Certified

VERIFICATION

Model Number: Saturn DigiSizer 5200


Type of Equipment: High Definition Digital Particle Size Analyzer


Applicable Directives: 89/336/EEC Electromagnetic Compatibility Directive
73/23/EEC Low Voltage Directive

Standards of Conformity: EN 61326 CISPR 16 and CISPR 16-1 Radiated and Conducted Emissions
IEC 61000-3-2 Power Frequency Harmonic Currents
IEC 61000-3-3 Voltage Fluctuation
IEC 61000-4-2, 4-3, 4-4, 4-5, 4-6, 4-8, and 4-88 Immunity Compliance
EN 61010-1 Part 1: General Safety Requirements

Other Tests: FCC Part 18 Industrial, Scientific & Medical Devices
Radiated and Conducted Emissions Limits

Micromeritics Instrument Corporation hereby declares that the instrument(s) specified above conform(s) to the Directive(s) and Standard(s), when installed and operated in accordance with the specifications set forth by Micromeritics. The original copy of this document is kept at Micromeritics with copies of the relevant test data and certificates, which constitute the required technical file for self declaration.


signature


date

Graham Killip

print name

Director of Engineering

position