



## CE DECLARATION OF CONFORMITY

Micromeritics Instrument Corporation, Inc.  
One Micromeritics Drive  
Norcross, Georgia 30093-1877 U.S.A.  
ISO 9001:2000 Certified

### VERIFICATION

**Model Number:** TriStar II 3020

**Type of Equipment:** Surface Area and Porosity Analyzer

**Applicable Directive:** **2004/108/EEC European Union (EU) Emissions and Immunity**  
**73/23/EEC Low Voltage Directive European Union (EU) Safety**

**Standards of Conformity:** **EN 61326: 1997 +A1: 1998 +A2:2001**  
IEC 61000-4-2 +A1+A2:2000 Electrostatic Discharge  
EN 61326 CISPR 11:2003 (RE) Radiated Emissions  
EN 61326 CISPR 11:2003 (CE) Conducted Emissions  
IEC 61000-4-3:2002 Radiated RF Electromagnetic Immunity  
IEC 61000-3-2:2000 Harmonic Current Emissions  
IEC 61000-3-3 +A1:2001 Voltage Fluctuations and Flicker  
IEC 61000-4-4:2004 Electrical Fast Transient/Burst  
IEC 61000-4-5 +A1:2000 Surge Immunity  
IEC 61000-4-6:2003 Conducted Immunity  
IEC 61000-4-11:2004 Voltage Dips, Interruptions, and Variations

**EN 61010-1 Part 1: General Safety Requirements**

Micromeritics Instrument Corporation hereby declares that the instrument(s) specified above conform(s) to the Directive(s) and Standard(s), when installed and operated in accordance with the specifications set forth by Micromeritics. The original of this document is kept at Micromeritics with copies of the relevant test data and certificates, which constitute the required technical file for self declaration.

Stefan Koch  
Signature

2-5-2008  
Date

Stefan Koch  
Print name

Director of Engineering  
Position